

## Equipment Prescribing Guidelines

code: Turner MOV 0627

Equipment: Stand Turner

Indications for Use	Indications for Non-Use	Further Action to be Taken	Alternatives
<ul style="list-style-type: none"> <li>• Service user can stand up with minimal assistance but needs support of knees and a hand hold.</li> <li>• Service user is unable to turn safely with a walking frame.</li> <li>• Service user must be able to stand on turner while being turned through 90°.</li> </ul>	<ul style="list-style-type: none"> <li>• If service user needs supporting on the turner by carer to stay standing.</li> <li>• If service user has spatial orientation problems that make them unsafe to stand when turner is moved.</li> <li>• If service user is able to walk short distances and turn through 90° with a walking frame.</li> </ul>	<ul style="list-style-type: none"> <li>• Consider hazard of having to adjust transfer surfaces before sitting the service user.</li> <li>• Consider if the service user is able to maintain standing on the turner throughout the transfer.</li> <li>• Consider service users ability to lift feet onto turner.</li> <li>• Consider moving and handling issues of positioning the turner and moving seating surfaces into position.</li> </ul>	<p>Contact I.L.C. to discuss alternative equipment and suppliers.</p> <ul style="list-style-type: none"> <li>• Stand assist hoist, alternative stand aids.</li> </ul>